



**All India Institute of Medical Sciences
Bilaspur, H.P. 174001
Procurement Section**

E-mail: storeofficer@aiimsbilaspur.edu.in



File No. **AIIMBL/235/2026-HOD(BPS)**

Date: **01/09/2026**

**NOTICE FOR INVITING QUOTATION SUPPLY OF ETO INCUBATOR AT AIIMS
BILASPUR.**

QUOTATION NOTICE

Sealed quotations are invited from registered stockiest/ distributors/ manufacturers having GST No. for Supply of **ETO Incubator** items as per the detailed item/specifications mentioned in **Annexure "1"**. Sealed Quotations in a single envelope duly super subscribed at the top of the envelope as "**Quotation No. AIIMBL/235/2026-HOD(BPS) for the Supply of ETO Incubator with end Date 21/09/2026**" containing both the "Technical Bid" and "Price Bid" (in two separate envelopes) may be submitted so as to reach on or before **21/09/2026** up to 05:00 P.M in **Procurement Section, 2nd floor, Administrative Block, AIIMS Bilaspur, Himachal Pradesh PIN- 174037**. Bidders need to quote for all the items. Late bids will not be considered. The bids shall be opened in the presence of duly constituted purchase committee for the purpose of technical evaluation; the bidder is required to submit the following documents.

A. Technical Bid (Envelope) Contains:

1. Profile-Name & Full Address of the firm & year of establishment.
2. Copy of permanent GST registration certificate.
3. Copies of supply orders secured during last 2 year for the similar products.
4. Turnover for last two years (2024-25 & 2025-26).
5. Self-declaration that the firm is not debarred by MoHFW or Department of Commerce.
6. Full description, make and quantity of the item must be mentioned in the quotation.
7. Technical Evaluation will be based on total value wise.

In case the Technical Committee rejects a sample (if asked for) or the bid is "Not Qualified" on technical grounds, the financial bid in respect of that item will not be considered.

Annexure- "1"

List of Items with Description is as follow

Sr No.	Item Name	Quantity
1	ETO INCUBATOR	01



**All India Institute of Medical Sciences
Bilaspur, H.P. 174001
Procurement Section**

E-mail: storeofficer@aiimsbilaspur.edu.in



सत्यमेव जयते

B. Price Bid: The Rates and total cost must be quoted in both words and figures (overwriting not allowed) in following format. Selection will be made purely on the basis of lowest price quoted by technically qualified firms. Validity of the quotation should be for a minimum period of 90 days in case discrepancy between unit price & total price, the unit price shall prevail.

Price Bid submission format for items with quantity required is as under: -

Sr. No.	Item Detail	Qty	Make	Model	HSN Code	Unit Price	GST Rate (%)	Unit Rate with GST	Total Amount	Amount In Figure
1	ETO INCUBATOR	01								

C. Other terms and conditions will be as follows: -

1. Analysis/ Batch report is submitted along with supply.
2. Supply will be F.O.R. at AIIMS-Bilaspur
3. Supply will have to be arranged within 15 days from the issue of supply order. Delayed supplies beyond 15 days from the date of Supply Order will be subject to LD @0.5% per week or part thereof, on the contract price subject to maximum of 10% of contract price beyond which the supply order will be liable to be cancelled.
4. Supply must conform to samples wherever asked for.
5. The Institute reserves the right to reject the goods if the same are not found in accordance with the specifications approved/ asked for. In case there is a short/defective supply, the firm will be informed and the defective material will be lifted from the concerned department/ Store by the supplier at their own cost within a two weeks period. The Institute will not bear any expenses on this account.
6. Samples will be preserved till final settlement.
7. Analysis/ Batch report is submitted along with supply.
8. Part supply will not be accepted.
9. No revision in rate (on higher side) will be accepted at any stage.
10. The firm shall not assign or sublet the work/job or any part of it to any other firm.
11. Billing will be in the name of Executive Director, AIIMS-Bilaspur. Payment will be made after the item has been received, inspected and accepted. No advance payment will be made at any stage.
12. Taxes at other government levies will be paid extra as applicable.
13. Venue for sample verification (if asked for) and quotation opening shall be in the **Procurement Section, AIIMS Bilaspur.**
14. In case the manufacturer participates, dealers will not be allowed or disqualified.
15. For any query, please contact Faculty In charge, E-mail storeofficer@aiimsbilaspur.edu.in




**All India Institute of Medical Sciences
Bilaspur, H.P. 174001
Procurement Section**

E-mail: storeofficer@aiimsbilaspur.edu.in



Check List for Technical Bid

S. No.	Document/Requirement Description	Attached (✓)
1.	Profile-Name & Full Address of the firm & year of establishment.	☐
2	Copy of permanent GST registration certificate.	☐
3.	Copies of supply orders secured during last 2 year for the similar products.	☐
4	Turnover for last two years (2024-25 & 2025-26).	☐
5	Self-declaration that the firm is not debarred by MoHFW or Department of Commerce.	☐
6	Full description, make and quantity of the item must be mentioned in the quotation.	☐
7	Technical Evaluation will be based on total value wise.	☐


Faculty In Charge
Procurement
AIIMS-Bilaspur H.P

Rapid
Incubator for
EO
Sterilization
Assurance

- Equipment should show remaining minutes left of BI incubation.
- Should be able to Configure Auto-reader remotely from a PC.
- Should have a built-in web app to allow create documentation that can be printed or saved electronically.
- Should have an option to connect to select sterile processing instrument tracking systems.
- Auto reader, self-calibrating, to read Rapid Attest BI, with ten incubation wells
- Positive BI Read-out time within 60 to 180 Mints by an audible alarm.
- It should be FDA Approved
- Should have IEC 61326-1 Electrical equipment for measurement, control, and laboratory use -EMC requirements - Part 1: General requirements
- Should Comply with the EMC requirements of the CE mark EMC Directive 2004/108/EC.
- It should be USFDA Approved , and Should have a BSI Kitemark Certificate.